

OZONE ANALYZER INTERCOMPARISON
 Thermo 49iPS_ACTRIS CM21267121

intercomparison place : Monte Cimone
 intercomparison operator: Paolo Cristofanelli, Francescopiero Calzolari
 Transfer Standard: Thermo 49iPs s/n: 1404860524
 Transfer Standard has been evaluated by ENPA on 1900-01-01 with SRP#15 giving
 slope of 1.000 and intercept of -0.3
 TS has been warmed-up for more then 12 hours and OA has been conditioning at
 200ppb for more then 2 hour
 OA has been evaluated at the following 9 concentration levels: 0, 15, 25, 50,
 75, 80, 100, 125 e 150 ppb

OA and TS condition:
 OA CM21267121 BKG=0.0 ; Coeff=1.017
 TS 1404860524 BKG=-0.3 ; Coeff=1.000
 intercomparison start : 2023-11-29 07:52:00 ; intercomparison end : 2023-11-29
 15:10:00
 LinregressResult(slope=0.9651200005999632, intercept=0.17526669640793102,
 rvalue=0.999988927333545, pvalue=3.4437424783629236e-55,
 stderr=0.0003294928407717181, intercept_stderr=0.028432579982436992)

Linear regression results OAm_{mean} = TS_{mean}*slope + intercept:
 TS Transfer Standard
 OA 03 Analyzer

slope	= 0.965120	slope_stderr	= 3.295e-04
intercept	= 0.175267	intercept_stderr	= 2.843e-02
rsquare	= 0.999998	covariance	= -4.202e-22

TS_{mean}: TS average [03] for each calibration step
 OA_{mean}: OA average [03] for each calibration step
 Predicted = OA_{mean}*slope + intercept
 TS_{std}: TS standard deviation [03] for each calibration step
 OA_{std}: OA standard deviation [03] for each calibration step
 Residual = TS - predicted
 Deviation = OA - TS

TS _{mean}	TS _{std}	OA _{mean}	OA _{std}	predicted	residual	deviation
-0.096	0.104	-0.165	0.186	0.016	-0.112	-0.069
74.729	0.138	77.240	0.270	74.721	0.007	2.511
149.721	0.095	154.886	0.150	149.659	0.062	5.165
124.702	0.102	129.234	0.184	124.901	-0.200	4.532
24.685	0.100	25.407	0.163	24.696	-0.011	0.722
99.682	0.105	103.147	0.225	99.724	-0.042	3.464
79.663	0.123	82.365	0.198	79.668	-0.005	2.703
49.722	0.078	51.320	0.182	49.705	0.017	1.597
14.940	0.482	15.232	0.544	14.876	0.064	0.292
99.682	0.105	103.013	0.178	99.595	0.087	3.331
149.720	0.113	154.879	0.285	149.653	0.067	5.160
0.039	0.125	-0.114	0.119	0.065	-0.026	-0.153
49.672	0.113	51.296	0.169	49.682	-0.009	1.623
24.723	0.087	25.306	0.172	24.598	0.124	0.583
99.714	0.103	103.205	0.259	99.781	-0.067	3.492
124.701	0.051	129.049	0.226	124.723	-0.022	4.348
14.709	0.110	15.031	0.156	14.682	0.027	0.322
74.697	0.098	77.164	0.229	74.648	0.049	2.467
99.724	0.113	103.080	0.317	99.660	0.064	3.356
79.703	0.078	82.457	0.157	79.756	-0.053	2.754
0.007	0.083	-0.151	0.106	0.030	-0.023	-0.158

U_{noise}: OA_{std} average = 0.213
 U_{linearity}: Residual standard deviation = 0.073
 U_{repeat} = sqrt(U_{noise}² + U_{linearity}²) = 0.225

$$\begin{aligned} \text{Udrift} &= \sqrt{0.58^2 + (0.0025 * C)^2} &= 0.632 \\ U &= \sqrt{U_{\text{repeat}}^2 + \text{Udrift}^2} &= 0.671 \\ C & &= 100.0 \end{aligned}$$

compensation equation to obtain unbiased concentration
 $[O_3]_{\text{unbiased}} = ([O_3] * 0.965) + 0.175$

Intercomparison 49iPS_ACTRIS s/n CM21267121 date : 20231129

