

OZONE ANALYZER INTERCOMPARISON  
 Thermo 49i CM08460046

intercomparison place : Bologna  
 intercomparison operator: Francescopiero Calzolari, Simonetta Montaguti  
 Transfer Standard: Thermo 49iPs s/n: 1404860524  
 Transfer Standard has been evaluated by ENPA on 1900-01-01 with SRP#15 giving  
 slope of 1.000 and intercept of -0.3  
 TS has been warmed-up for more then 12 hours and OA has been conditioning at  
 200ppb for more then 2 hour  
 OA has been evaluated at the following 9 concentration levels: 0, 15, 25, 50,  
 75, 80, 100, 125 e 150 ppb

OA and TS condition:  
 OA CM08460046 BKG=-0.2 ; Coeff=1.020  
 TS 1404860524 BKG=-0.3 ; Coeff=1.000  
 intercomparison start : 2024-01-19 07:24:00 ; intercomparison end : 2024-01-19  
 14:42:00  
 LinregressResult(slope=1.012692380195508, intercept=-0.5830908734453146,  
 rvalue=0.999977739628054, pvalue=8.282419663499854e-43,  
 stderr=0.0015502052332417463, intercept\_stderr=0.12843431396927532)

Linear regression results OAmean = TSmean\*slope + intercept:  
 TS Transfer Standard  
 OA O3 Analyzer

slope	= 1.012692	slope_stderr	= 1.550e-03
intercept	= -0.583091	intercept_stderr	= 1.284e-01
rsquare	= 0.999955	covariance	= 1.608e-20

TSmean: TS average [O3] for each calibration step  
 OAmean: OA average [O3] for each calibration step  
 Predicted = OAmean\*slope + intercept  
 TSstd: TS standard deviation [O3] for each calibration step  
 OAstd: OA standard deviation [O3] for each calibration step  
 Residual = TS - predicted  
 Deviation = OA - TS

TSmean	TSstd	OAmean	OAstd	predicted	residual	deviation
-0.026	0.117	0.900	0.134	0.328	-0.354	0.926
74.714	0.066	73.340	0.220	73.688	1.026	-1.374
149.697	0.096	148.030	0.142	149.326	0.372	-1.667
124.677	0.084	123.620	0.125	124.606	0.071	-1.057
24.702	0.123	24.770	0.195	24.501	0.200	0.068
99.726	0.097	98.950	0.180	99.623	0.103	-0.776
79.730	0.077	79.200	0.173	79.622	0.108	-0.530
49.699	0.107	49.540	0.180	49.586	0.114	-0.159
14.680	0.110	14.930	0.127	14.536	0.144	0.250
99.693	0.138	98.970	0.155	99.643	0.050	-0.723
149.718	0.069	148.820	0.178	150.126	-0.408	-0.898
0.040	0.099	0.900	0.148	0.328	-0.289	0.860
49.718	0.107	49.450	0.120	49.495	0.224	-0.268
24.675	0.096	24.820	0.166	24.552	0.123	0.145
99.710	0.059	99.200	0.155	99.876	-0.166	-0.510
124.690	0.122	124.120	0.108	125.112	-0.422	-0.570
14.660	0.077	15.140	0.080	14.749	-0.089	0.480
74.708	0.101	74.430	0.142	74.792	-0.083	-0.278
99.716	0.071	99.340	0.102	100.018	-0.302	-0.376
79.703	0.107	79.550	0.206	79.977	-0.273	-0.153
0.111	0.115	0.830	0.110	0.257	-0.147	0.719

Unoise: OAstd average = 0.150  
 Ulinearity: Residual standard deviation = 0.327  
 Urepeat = sqrt(Unoise^2 + Ulinearity^2) = 0.360

$$\begin{aligned} \text{Udrift} &= \sqrt{0.58^2 + (0.0025 * C)^2} &= 0.632 \\ U &= \sqrt{U_{\text{repeat}}^2 + \text{Udrift}^2} &= 0.727 \\ C & &= 100.0 \end{aligned}$$

compensation equation to obtain unbiased concentration  
 $[O_3]_{\text{unbiased}} = ([O_3] * 1.013) - 0.583$

# Intercomparison 49i s/n CM08460046 date : 20240119

