

OZONE ANALYZER INTERCOMPARISON
Thermo 49c 04275089232

intercomparison place : Lab 103 ISAc Bologna
intercomparison operator: Maurizio Busetto, Paolo Cristofanelli
Transfer Standard: Thermo 49iPs s/n: CM21267121
Transfer Standard has been evaluated by NaN on NaN with SRP#15 giving slope of NaN
and intercept of NaN
TS has not been warmed-up for more then 12 hours and OA has not been conditioning
at 200ppb for more then 2 hour
OA has been evaluated at the following 5 concentration levels: 0, 25, 50, 75, 100 e
125 ppb

OA and TS condition:

OA 04275089232 BKG=-0.2 ; Coeff=1.010

TS CM21267121 BKG=-0.0 ; Coeff=1.017

intercomparison start : 2023-06-14 08:45:00 ; intercomparison end : 2023-06-14
15:52:00

LinregressResult(slope=1.0201375110314574, intercept=0.09247082619982194,
rvalue=0.9999933495122809, pvalue=6.8129087474366375e-43,
stderr=0.0009023565359760328, intercept_stderr=0.047008086053586216)

risultati regressione lineare OAmean = TSmean*slope + intercept:

TS Transfer Standard

OA 03 Analyzer

slope	= 1.020138	slope_stderr	= 9.024e-04
intercept	= 0.092471	intercept_stderr	= 4.701e-02
rsquare	= 0.999987	covariance	= 1.464e-21

TSmean : media [03] TS ogni step di calibrazione

OAmean : media [03] OA per ogni step di calibrazione. NB: OAmean Ã" ottenuta dalle misure dell'analizzatore riportate a slope=1.0 e bkg=0.0

Predicted = TSmean*slope + intercept

TSstd : standard deviation [03] TS per ogni step di calibrazione

OAstd : standard deviation [03] OA per ogni step di calibrazione

Residual = TS - predicted

Deviation = OA - TS

TSmean	TSstd	OAmean	OAstd	predicted	residual	deviation
0.024	0.129	-0.128	0.098	-0.038	0.062	-0.152
14.996	0.147	14.533	0.122	14.918	0.078	-0.463
50.000	0.096	48.916	0.180	49.994	0.006	-1.084
24.974	0.178	24.369	0.167	24.952	0.022	-0.605
75.020	0.169	73.367	0.150	74.937	0.083	-1.653
100.011	0.102	97.786	0.226	99.848	0.163	-2.225
0.078	0.127	-0.149	0.077	-0.060	0.138	-0.227
25.026	0.201	24.405	0.154	24.989	0.037	-0.621
75.011	0.092	73.521	0.138	75.094	-0.083	-1.490
99.989	0.090	98.075	0.166	100.142	-0.153	-1.914
14.974	0.091	14.562	0.146	14.948	0.026	-0.412
50.022	0.194	48.886	0.277	49.963	0.059	-1.136
0.098	0.132	-0.133	0.111	-0.043	0.142	-0.231
49.985	0.119	48.907	0.195	49.984	0.001	-1.078
100.000	0.134	97.943	0.166	100.008	-0.008	-2.057
24.999	0.190	24.531	0.220	25.117	-0.118	-0.468
14.993	0.216	14.586	0.208	14.972	0.021	-0.407
49.995	0.087	48.989	0.160	50.068	-0.073	-1.006
-0.010	0.155	0.293	1.285	0.392	-0.402	0.304

Unoise : media OAstd	= 0.224
Ulinearity : standard deviation Residual	= 0.129
Urepeat = sqrt(Unoise^2 + Ulinearity^2)	= 0.258
Udrift = sqrt(0.58^2+(0.0025*C)^2)	= 0.632
U = sqrt(Urepeat^2+Udrift^2)	= 0.682
C	= 100.0

compensation equation to obtain unbiased concentration
[O3unbiased]=([OA]*1.020)+0.092

Intercomparison 49c s/n 04275089232 date : 20230614

